

# Certificate of Traceability

## Level I Laboratory Certification

This certifies that Covance, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:


Instrument: <b>Variant II Turbo 2.0</b>	Calibrator Lot: <b>64287389, 64287390</b>	Column Lot: <b>70234AM</b>
Reagent Lot: <b>64307603, 64317060, 64309775</b>	Calibrator Assigned Values: <b>5.2%, 10.2%</b>	

Date of Certification: May 1, 2020

Certification Expires: May 1, 2021

  
NGSP Steering Committee Chair

  
NGSP Network Coordinator

  
SRL director/ supervisor